

This Page Is Inserted by IFW Operations
and is not a part of the Official Record

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images may include (but are not limited to):

- BLACK BORDERS
- TEXT CUT OFF AT TOP, BOTTOM OR SIDES
- FADED TEXT
- ILLEGIBLE TEXT
- SKEWED/SLANTED IMAGES
- COLORED PHOTOS
- BLACK OR VERY BLACK AND WHITE DARK PHOTOS
- GRAY SCALE DOCUMENTS

IMAGES ARE BEST AVAILABLE COPY.

**As rescanning documents *will not* correct images,
please do not report the images to the
Image Problem Mailbox.**

Medical Imaging III: Image Processing

Roger H. Schneider
Samuel J. Dwyer III
R. Gilbert Jost
Chairs/Editors

31 January–3 February 1989
Newport Beach, California

Sponsored by
SPIE—The International Society for Optical Engineering

In Cooperation with
American Association of Physicists in Medicine
American College of Radiology
Center for Devices and Radiological Health, FDA
Institute for Regulatory Science



Volume 1092

Fourier method for correction of depth-dependent collimator blurring

Robert M. Lewitt, Paul R. Edholm* and Weishi Xia

Department of Radiology, University of Pennsylvania
Blockley Hall, 418 Service Drive, Philadelphia, Pennsylvania 19104

*Department of Diagnostic Radiology, University of Linköping, Sweden

ABSTRACT

A method is described for preprocessing projection data prior to image reconstruction in single-photon emission computed tomography. The projection data of the desired spatial distribution of emission activity is blurred by the point-response function of the collimator that is used to define the range of directions of gamma-ray photons reaching the detector. The point-response function of the collimator is not stationary, but depends on the distance from the collimator to the point. Conventional methods for deblurring collimator projection data are based on approximating the actual depth-dependent point-response function by a spatially-invariant blurring function, so that deconvolution methods can be applied independently to the data at each angle of view. The method described in this paper is based on Fourier analysis of the multi-angular data set as a whole, using special depth-dependent characteristics of the Fourier coefficients to achieve spatially-variant inverse filtering of the data in all views simultaneously. Preliminary results are presented for simulated data with a simple collimator model.

1. INTRODUCTION

Nuclear medicine involves the determination of functional and metabolic information about selected human organ systems from the distribution of radioactive tracers administered to the patient. The conventional instrument for imaging the distribution of activity combines a gamma camera to detect and map the received radiation and a collimator to limit the range of directions of the detected photons. A stationary gamma camera with its associated collimator produces a projection image of the activity distribution within its field of view. This type of image has many uses in nuclear medicine, but it has the disadvantage that overlying and underlying parts of the activity distribution appear superimposed, analogous to the superposition problem inherent in the conventional x-ray projection image. Single-photon emission computed tomography (SPECT)¹ is a method for eliminating the superposition problem in nuclear medicine imaging. SPECT produces images of cross-sectional slices free from out-of-slice contributions by computer processing of the set of projection data that is acquired by a collimated gamma camera for many angles of view as it rotates around the patient.

Although SPECT has now found widespread use in clinical nuclear medicine², there remain some significant technical challenges in the design of the instrumentation and in the processing of the data. The design of the collimator involves an inherent compromise between resolution and sensitivity. As the collimator holes are made narrower and/or longer, the resolution is improved, but fewer of the emitted photons are accepted by the collimator. Also, the acceptance angle of a collimator hole relative to a point source depends both on the distance from the point to the collimator and on the distance from the point to the line formed by extending the central axis of the collimator hole into the activity distribution³. The point-response function of the collimator and the consequent blurring in the detected projection of a point source are therefore depth-dependent. As the collimator is rotated around the patient, a fixed point source that is offset from the center of rotation will contribute greater or lesser amounts to the projection data as the distance from the point to the collimator varies. Our approach to the problem of depth-dependent collimator blurring is the main subject of this paper.

In addition to the problem of position-variance of the collimator's point-response function, there are other significant problems that limit the quantitative capabilities of SPECT⁴. Because the radiation dose to the patient must be minimized, and because the collimator accepts only a very small fraction of the emitted photons, the number of photons contributing to each measurement is relatively small, resulting in large statistical uncertainties in the measured data. Other significant problems that affect the quality of the data are caused by photon attenuation and photon scattering between the original point of emission and the detector.

In this paper we describe the basic principles of our approach to the correction of depth-dependent collimator blurring in SPECT. We use a simplified model of the actual image-formation process in order to concentrate on the basic ideas. We consider a two-dimensional (2-D) analog of the actual 3-D situation: specifically, we consider a 2-D distribution of activity in the plane, with 1-D projection data collected by a 1-D collimator for each angle of view around the activity distribution. The present formulation of our approach does not include photon attenuation or scattering.

The conventional approach to the problem of collimator blurring assumes that the depth-dependent point-response function can be approximated by a spatially-invariant blurring function, so that inverse filtering can be applied to the projection data in each view independently⁵. Our approach involves simultaneous inverse filtering of the data in all views prior to conventional image reconstruction from projections, and is an extension of our earlier work⁶ on properties of the Fourier decomposition of the sinogram. Ogawa *et al.*⁷ have recently presented a method for correction of depth-dependent collimator blurring that also involves Fourier decomposition of the sinogram, but their formulation of the problem and the coefficients of their filter function are different from ours. Fourier analysis of the sinogram has also been used by Hawkins *et al.*⁸ as the basis of an alternative reconstruction method for SPECT. Their formulation is directed toward solving the problem of photon attenuation (which is assumed to be uniform within a convex boundary), rather than the problem of depth-dependent collimator blurring which is the motivation for the preprocessing method in this paper.

2. NOTATION AND PRELIMINARIES

2.1 Coordinates and projections

Our notation for object and detector coordinates in the 2-D plane is shown in Fig. 1. In object coordinates, (x, y) represents a point in the plane, and $f(x, y)$ denotes the activity (emission density) at (x, y) . The detector coordinate axes rotate with the collimator and detector about the origin as shown in Fig. 1, where the detector coordinates (s, t) are rotated by θ from the fixed object coordinates (x, y) . We denote by $f^\theta(s, t)$ the activity at (s, t) in rotated (detector) coordinates:

$$f^\theta(s, t) = f(s \cos \theta - t \sin \theta, s \sin \theta + t \cos \theta) \quad (1)$$

A line in the plane parallel to the t axis is specified by its pair of parameters (s, θ) . We denote by $p(s, \theta)$ the value of the line integral of activity along the line (s, θ) :

$$p(s, \theta) = \int f^\theta(s, t) dt \quad (2)$$

The inversion process necessary to estimate f from p is known as image reconstruction from projections^{9, 10}.

2.2 Collimator response function

The assumed collimator consists of a row of parallel holes, where the long axis of each hole is parallel to the t axis, and the row of holes is at a fixed distance t_c from the center of rotation. In our 2-D model, these "holes" are actually 1-D bins separated by partitions (septa). For a collimator hole centered on the line (s, θ) , as shown in Fig. 2, the detected response to a point source of activity located at (s', t) in detector coordinates depends on t (the distance along the hole axis minus t_c) and on the distance $s' - s$ from the point to the center-line of the hole. We denote this point-response function by h , and we assume that the point response for a fixed t is symmetric about the center-line, so that

$$h(s - s', t) = h(s' - s, t) = h(|s' - s|, t) \quad (3)$$

We denote by $p_c(s, \theta)$ the value of the "fan integral" of activity for the collimator hole centered on the ray (s, θ) :

$$p_c(s, \theta) = \int \left[\int f^\theta(s', t) h(s - s', t) ds' \right] dt \quad (4)$$

In the following sections of this paper, we present a method to estimate p from p_c so that conventional image reconstruction algorithms for p can be used for collimator projection data p_c that has been preprocessed using this method.

2.3 Point source geometry

Consider a point source of activity at polar coordinates r_p and ϕ_p , as shown in Fig. 3. The (x, y) coordinates of this point are $(r_p \cos \phi_p, r_p \sin \phi_p)$ and the (s, t) coordinates are denoted by $(s_p(\theta), t_p(\theta))$, where

$$s_p(\theta) = r_p \cos(\phi_p - \theta) \quad \text{and} \quad t_p(\theta) = r_p \sin(\phi_p - \theta) \quad (5)$$

The collimator response to this point source is

$$p_c(s, \theta) = h(s - s_p(\theta), t_p(\theta)) \quad (6)$$

2.4 Fourier analysis of point-response function

The 1-D Fourier transform of $h(s, t)$ with respect to its first variable is denoted by $\hat{h}(\omega, t)$, where

$$\hat{h}(\omega, t) = \int_{-\infty}^{+\infty} h(s, t) e^{-i\omega s} ds \quad (7)$$

For the shifted point-response function, we find

$$\hat{h}(\omega, t) e^{-i\omega s'} = \int_{-\infty}^{+\infty} h(s - s', t) e^{-i\omega s} ds \quad (8)$$

2.5 Fourier analysis of the sinogram

The sinogram is the 2-D map of p as a function of s and θ . We make use of 2-D Fourier analysis of the sinogram, which is periodic in θ (but not in s). The Fourier coefficients are denoted by $\hat{p}(\omega, n)$, where ω is a real number and n is an integer, and we express the Fourier pair as

$$\hat{p}(\omega, n) = \frac{1}{2\pi} \int_0^{2\pi} \int_{-\infty}^{+\infty} p(s, \theta) e^{-i(\omega s + n\theta)} ds d\theta \quad (9)$$

$$p(s, \theta) = \frac{1}{2\pi} \int_{-\infty}^{+\infty} \sum_{n=-\infty}^{\infty} \hat{p}(\omega, n) e^{i(\omega s + n\theta)} d\omega \quad (10)$$

3. FOURIER ANALYSIS FOR A POINT SOURCE

3.1 Exact analysis

The Fourier coefficients of the collimated projections are defined by

$$\hat{p}_c(\omega, n) = \frac{1}{2\pi} \int_0^{2\pi} \int_{-\infty}^{+\infty} p_c(s, \theta) e^{-i(\omega s + n\theta)} ds d\theta \quad (11)$$

For the case of a point source at $(r_p \cos \phi_p, r_p \sin \phi_p)$, we substitute $h(s - s_p(\theta), t_p(\theta))$ for $p_c(s, \theta)$, then integrate over s to obtain

$$\hat{p}_c(\omega, n) = \frac{1}{2\pi} \int_0^{2\pi} \hat{h}(\omega, t_p(\theta)) e^{-i[\omega r_p \cos(\phi_p - \theta) + n\theta]} d\theta \quad (12)$$

3.2 Asymptotic analysis

We now examine the characteristics of the complex exponential factor in the integrand, and indicate how the principle of stationary phase¹¹ may be used to obtain approximations to the integral. Our purpose here is simply to outline the steps involved, without attempting to present a mathematically rigorous derivation. We introduce the notation $\Phi(\theta)$, where

$$\Phi(\theta) = \omega r_p \cos(\phi_p - \theta) + n\theta = \omega s_p(\theta) + n\theta \quad (13)$$

so that

$$\hat{p}_c(\omega, n) = \frac{1}{2\pi} \int_0^{2\pi} \hat{h}(\omega, t_p(\theta)) e^{-i\Phi(\theta)} d\theta \quad (14)$$

Figure 4 illustrates the behavior of Φ and $\cos \Phi$ as functions of θ for $(\omega, n) = (-80, 40)$. For ω large, $e^{-i\Phi(\theta)}$ oscillates rapidly as a function of θ , except near angles (denoted by θ_1 and θ_2) at which $\partial \Phi / \partial \theta = 0$. The relatively small portions of the range of integration centered around the angles θ_1 and θ_2 are the only ones that make significant contributions to the

value of the integral, since the integral over the oscillatory parts of the function is almost zero. If $\hat{h}(\omega, t_p(\theta))$ varies only slowly as a function of θ , compared with $e^{-i\Phi(\theta)}$, then $\int_0^{2\pi}$ can be approximated by $\int_{\theta_1-\epsilon}^{\theta_1+\epsilon} + \int_{\theta_2-\epsilon}^{\theta_2+\epsilon}$ as follows:

$$\hat{p}_c(\omega, n) \approx \frac{1}{2\pi} \int_{\theta_1-\epsilon}^{\theta_1+\epsilon} \hat{h}(\omega, t_p(\theta)) e^{-i\Phi(\theta)} d\theta + \frac{1}{2\pi} \int_{\theta_2-\epsilon}^{\theta_2+\epsilon} \hat{h}(\omega, t_p(\theta)) e^{-i\Phi(\theta)} d\theta \quad (15)$$

We now take \hat{h} outside each of the integrals, since the intervals are small, and \hat{h} is assumed to be slowly varying, so that

$$\hat{p}_c(\omega, n) \approx \frac{1}{2\pi} \hat{h}(\omega, t_p(\theta_1)) \int_{\theta_1-\epsilon}^{\theta_1+\epsilon} e^{-i\Phi(\theta)} d\theta + \frac{1}{2\pi} \hat{h}(\omega, t_p(\theta_2)) \int_{\theta_2-\epsilon}^{\theta_2+\epsilon} e^{-i\Phi(\theta)} d\theta \quad (16)$$

We find θ_1 and θ_2 from the condition for stationary phase (i.e., $\partial\Phi/\partial\theta = 0$), where

$$\frac{\partial\Phi}{\partial\theta} = \omega r_p \sin(\phi_p - \theta) + n = \omega t_p(\theta) + n \quad (17)$$

$$\frac{\partial\Phi}{\partial\theta} = 0 \quad \text{leads to} \quad \sin(\phi_p - \theta) = -\frac{n}{\omega r_p} \quad \text{and} \quad t_p(\theta) = -\frac{n}{\omega} \quad (18)$$

For $|n| \leq |\omega r_p|$, there exist angles θ_1, θ_2 in the range $[0, 2\pi]$ for which $\partial\Phi/\partial\theta = 0$. We also require $t_p(\theta_1)$ and $t_p(\theta_2)$ in the expression for $\hat{p}_c(\omega, n)$, and the condition for stationary phase indicates directly that $t_p(\theta_1)$ and $t_p(\theta_2)$ have the remarkably simple form

$$t_p(\theta_1) = t_p(\theta_2) = -\frac{n}{\omega} \quad (19)$$

so that

$$\hat{h}(\omega, t_p(\theta_1)) = \hat{h}(\omega, t_p(\theta_2)) = \hat{h}\left(\omega, -\frac{n}{\omega}\right) \quad (20)$$

and

$$\hat{p}_c(\omega, n) \approx \hat{h}\left(\omega, -\frac{n}{\omega}\right) \left\{ \frac{1}{2\pi} \int_{\theta_1-\epsilon}^{\theta_1+\epsilon} e^{-i\Phi(\theta)} d\theta + \frac{1}{2\pi} \int_{\theta_2-\epsilon}^{\theta_2+\epsilon} e^{-i\Phi(\theta)} d\theta \right\} \quad (21)$$

For the ideal case of the infinite-resolution collimator, $\hat{p}_c \rightarrow \hat{p}$ and $\hat{h} \rightarrow 1$, so that $\hat{p}(\omega, n) \approx \{ \int + \int \}$ in the previous equation, and

$$\hat{p}_c(\omega, n) \approx \hat{h}\left(\omega, -\frac{n}{\omega}\right) \hat{p}(\omega, n) \quad (22)$$

This result shows that an estimate of $\hat{p}(\omega, n)$ can be recovered, in principle, from $\hat{p}_c(\omega, n)$ by

$$\hat{p}(\omega, n) \approx [\hat{h}(\omega, -n/\omega)]^{-1} \hat{p}_c(\omega, n) \quad (23)$$

where $[\hat{h}(\omega, -n/\omega)]^{-1}$ denotes the inverse filter.

The above Fourier analysis is based on the detected response to an arbitrary point source of activity. We find, however, that the final expressions for \hat{p}_c and \hat{p} are independent of the coordinates of the point source. By the principle of superposition, and the linearity of the operations involved, these expressions are therefore applicable to the general case of a non-localized distribution of activity.

4. PROPERTIES OF THE FOURIER COEFFICIENTS

4.1 Double-wedge characteristic

The analysis of the Fourier coefficients of the sinogram in the previous section shows that, for a point source of activity at polar coordinates r_p and ϕ_p , the condition for stationary phase is satisfied at two angles within the range of integration, provided $|n| \leq |\omega r_p|$. Assuming r_p non-negative, we have $|n/\omega| \leq r_p$. Let R_p be the largest value of r_p for any point in the activity distribution. Then R_p is the largest value of $|n/\omega|$ for which the stationary condition is satisfied for at least one point source in the activity distribution. For $|n/\omega| > R_p$, the stationary-phase condition is satisfied only for points outside the activity distribution and, as a consequence, the Fourier coefficients are almost zero. Because of this property, the Fourier coefficients form a double-wedge or "bow-tie" pattern in the (ω, n) plane, as shown in Fig. 5, which is described below. A more detailed discussion of the geometrical basis of this property of the Fourier coefficients is given in our earlier report⁶, and we note in passing that the coefficients of other projection expansions involving angular harmonics are found to have their own double-wedge properties, known as projection consistency conditions^{8,9}.

Figure 5 shows the results of experiments using a phantom consisting of two uniform disks of emission activity inside a unit circle, where the disks each have a radius of 0.1 and are centered at coordinates (0.8, 0) and (-0.8, 0). Figure 5a shows the sinogram of the simulated parallel projections $p(s, \theta)$ from an ideal detector, the image reconstructed from this sinogram, and the map of the calculated Fourier coefficients $\hat{p}(\omega, n)$. The double-wedge characteristic is evident. Projection data were also simulated for the collimator case, where the point-response function was calculated geometrically using the angle subtended at the point by the collimator aperture. Figure 5b shows the sinogram of the simulated collimator data, the image reconstructed from the collimator data, and the map of the Fourier coefficients of the collimator data.

4.2 Connection with Bessel functions

In the case of projections of a point source obtained by an infinite-resolution collimator, we find that the integral for the Fourier coefficients reduces to the integral representation for the Bessel function. Since the properties of Bessel functions have been the subject of extensive investigations¹¹, many interesting properties of this integral, and hence properties of the Fourier coefficients of ideal data, can be deduced from them.

The Bessel function of the first kind, order n , is denoted by J_n , and its integral representation is

$$J_n(z) = \frac{1}{2\pi} \int_{\alpha}^{2\pi+\alpha} e^{i(n\theta - z \sin \theta)} d\theta \quad (24)$$

where α is an arbitrary angle. For the ideal case of the infinite-resolution collimator, $h \rightarrow 1$, so that the projections of a point source at polar coordinates r_p and ϕ_p lead to

$$\hat{p}(\omega, n) = \frac{1}{2\pi} \int_0^{2\pi} e^{-i[\omega r_p \cos(\phi_p - \theta) + n\theta]} d\theta \quad (25)$$

In terms of Bessel functions, $\hat{p}(\omega, n)$ may be expressed as

$$\hat{p}(\omega, n) = J_n(\omega r_p) e^{-in(\phi_p + \pi/2)} \quad (26)$$

For non-zero n , the Bessel function $J_n(z)$ has a different behavior for $|z| < |n|$ than it has for $|z| > |n|$. If $n \geq 1$, then for increasing values of the argument z , beginning from $z = n$, the function first increases to its maximum value and then oscillates about the zero axis, similar to a decaying sinusoid, with a slowly decreasing amplitude of oscillation asymptotically proportional to $z^{-1/2}$. For decreasing values of z , beginning from $z = n$, the integral that represents the Bessel function does not include a point of stationary phase, and the function decreases rapidly and monotonically toward the zero axis, which it eventually intersects at $z = 0$. These characteristics of the Bessel function, together with

$$|\hat{p}(\omega, n)| = |J_n(\omega r_p)| \quad (27)$$

imply that the non-negligible values of $\hat{p}(\omega, n)$ are confined to a double-wedge shaped region in the (ω, n) map. The two lines $n = \pm \omega r_p$ mark the boundary of the double-wedge region. Considering a general object as a superposition of points, we still get the double-wedge shaped \hat{p} distribution as a result of the superposition of the transforms of the individual point objects.

The double-wedge shaped pattern in the (ω, n) plane due to each point object has a transition region near the boundary lines $n = \pm\omega r_p$. We now point out some relevant characteristics of the Bessel function in this transition region (see also Natterer¹², pages 197–199). Let Z_n denote the location of the first local maximum (this is also the global maximum) of $J_n(z)$. For n positive, and reasonably large (e.g., $n > 10$) the following large-order asymptotic approximations are useful (see Abramowitz and Stegun¹³, formulas 9.5.16, 9.5.20, 9.3.5 and 6.1.13):

$$Z_n \sim n + 0.809n^{\frac{1}{3}} \quad \text{with} \quad J_n(Z_n) \sim 0.674n^{-\frac{1}{3}} \quad \text{and} \quad J_n(n) \sim 0.447n^{-\frac{1}{3}} \quad (28)$$

For $z < n$, it is of interest to determine the value of z at which $J_n(z)$ is negligible compared to its maximum value. Numerically solving $J_n(z) = 0.01J_n(Z_n)$, we find that $z \sim n - 1.35n^{\frac{1}{3}}$. This value of z may be considered to mark the low-valued edge of the transition region. The width of the transition region is therefore dependent only weakly on n .

4.3 Frequency-distance relation

We first state in general terms a property of the Fourier coefficients that we call the “frequency-distance relation”⁶. We then show the implications of this property by means of an illustrative example. In section 3.2 we found that

$$t_p(\theta_1) = t_p(\theta_2) = -\frac{n}{\omega} \quad (29)$$

This remarkable result is what we refer to as the *frequency-distance relation for the Fourier coefficients*. It means that as θ goes from 0 to 2π , the most significant contribution of an arbitrary point object to the value of the coefficient $\hat{p}(\omega, n)$ occurs at an angle value of θ such that the point's distance t along the ray is $-n/\omega$. The importance of this relation lies in the possibility of modifying the coefficients within a narrow area around the line $n = -\omega t$ in the Fourier domain and obtaining the effect of modifying contributions to the projection data that have been emitted at a specific distance t from the detector. This is a new concept in projection imaging, initially presented in our previous report⁶, and it is not easy to present and to understand in the form of the general statement that we have just given. The discussion and specific examples that follow are intended to make clearer the meaning and implications of the frequency-distance relation for the Fourier coefficients.

Since t in the detector coordinates represents a signed distance from the object location to the s axis, a negative value of t means that the point of interest is located in the half of the object which is closer to the detector. When t is of positive value the half of the object which is farther away from the detector is indicated. From the frequency-distance relation $t_p = -n/\omega$, we find that in the (ω, n) plane, the Fourier coefficients \hat{p} in the two quadrants where $n/\omega > 0$ correspond to contributions to the projection data from point sources that are nearer to the detector than the center of rotation (where $t = 0$). The other two quadrants corresponding to $n/\omega < 0$ represent contributions to the projection data from point sources that are farther from the detector than the center of rotation (Fig. 6).

Assuming data are collected for a full 360° range of views, a particular point source in the object will have negative t (i.e., it will be nearer to the detector than is the center of rotation) for a subinterval of 180° within the full range of views. This subinterval of 180° is sufficient for accurate reconstruction of this point. Different point sources will have different subintervals, each of length 180° , located within the full range of views and corresponding to negative t in detector coordinates. The power of the Fourier analysis involving the frequency-distance relation is that it enables the full-range projection data to be decomposed into another data set where *each point* is represented by its own appropriate subinterval within the full range of views.

A simple experiment illustrates this property. We see from the analysis above that if only the coefficients for which $n/\omega > 0$ were selected, then the data used for reconstruction would involve only contributions from points near to the detector, where the resolution and sensitivity are best. The experiment involves the following steps. First, the emission projections $p_c(s, \theta)$ are obtained from the collimated detector model for a complete rotation in θ . Then, the 2-D discrete Fourier transform of $p_c(s, \theta)$ produces $\hat{p}_c(\omega, n)$. Now mask out the \hat{p}_c data from the second and fourth quadrants. Finally, the 2-D inverse Fourier transform of the selected coefficients in the first and third quadrants produces a modified sinogram $\tilde{p}_c(s, \theta)$. Reconstruction is done from these selected components of the projection data to get an improved image.

Figure 7a shows the blurred sinogram from the simulated SPECT detector, its Fourier decomposition and the image of two blurred disks reconstructed from it. From Fig. 7b we can see the improvement in the reconstruction after following the procedure mentioned above. In comparison, Fig. 7c shows what happens if we do the complementary procedure, selecting those quadrants that involve contributions to the projection data from point sources that are farther from the detector than is the center of rotation. The collimator resolution is poor for such point sources, as Fig. 7c shows. This experiment shows that heavily blurred data can be simply picked out and discarded from collimated projections by processing in the Fourier domain. The frequency-distance relation is more powerful than this, however, because it can be used to correct each blurred projection depth-by-depth along the projection ray without discarding any detected events, as we describe in the next section.

5. DEPTH-DEPENDENT SPATIAL FILTERING OF THE SINOGRAM

5.1 Procedure for filtering of projection data

The analysis in section 3.2 leads to the following Fourier method for correction of depth-dependent collimator blurring:

- 1) Compute $\hat{p}_c(\omega, n)$ by the 2-D discrete Fourier transform of the data $p_c(s, \theta)$ obtained for a complete rotation around the patient;
- 2) Recover estimate of $\hat{p}(\omega, n)$ from $\hat{p}_c(\omega, n)$ by $\hat{p}(\omega, n) \approx [\hat{h}(\omega, -n/\omega)]^{-1} \hat{p}_c(\omega, n)$, where the inverse filter is determined by the procedure described in section 5.2;
- 3) Compute estimate of $p(s, \theta)$ by 2-D inverse Fourier transform.

5.2 Inverse filter design

The first step in computing the inverse filter is to specify the point-response function $h(s, t)$ of the collimator-detector combination. This combined response function has two components, namely, the geometrical response function of the collimator and the intrinsic point-response function of the detector.

In experimental measurements of point-response functions, it has been found¹⁴ that a Gaussian function in s provides a good fit to the experimental data for each t . With a Gaussian representation for $h(s, t)$, its 1-D Fourier transform $\hat{h}(\omega, t)$ can be found analytically for each t , so that $\hat{h}(\omega, -n/\omega)$ can be evaluated at the required sample points in the (ω, n) space.

Once $\hat{h}(\omega, -n/\omega)$ is available, an inverse filter (which we denote by $[\hat{h}(\omega, -n/\omega)]^{-1}$) may be designed. The simple reciprocal of $\hat{h}(\omega, n)$ is a reasonable inverse filter only for those (ω, n) where the signal-to-noise ratio is high and $|\hat{h}(\omega, n)|$ is not too small. At high frequencies, the signal-to-noise ratio is low, and $|\hat{h}(\omega, n)|$ is small, so that the high-frequency characteristic of a useful inverse filter must deviate from the simple reciprocal in order to avoid excessive noise amplification. An appropriate high-frequency characteristic is also necessary in conventional data preprocessing in SPECT^{5, 15} for the same reason. The Wiener filter and Metz filter have been used¹⁶⁻¹⁹ to construct deblurring filters for SPECT projection data based on the assumption that the actual blurring can be approximated by a uniform blurring corresponding to the center of rotation. We are presently investigating the properties of depth-dependent inverse filters derived from Wiener, Metz, and other filter prototypes.

5.3 Results

Figure 8 shows the results of a preliminary simulation study to test the basic principles of the method. The phantom that was simulated consists of circular regions of various sizes, each having a uniform distribution of activity. Collimator data $p_c(s, \theta)$ was simulated for 135 samples in s , 360 samples in θ (complete rotation) for a parallel-hole collimator having a hole depth/diameter ratio of 5. In this preliminary study, the data were noiseless (infinite statistics) with no attenuation or scattering. The geometrical point-response function of the collimator was used in the simulation of the data, and a Gaussian approximation was used for the purpose of calculating the inverse filter, as described in section 5.2. Although this simulation represents an idealized imaging situation, the result does indicate that the method has potential for deblurring projection data, and hence for improving the fidelity of the reconstructed image.

6. CONCLUSIONS

We conclude that a linear filtering approach based on Fourier transforms has potential for the correction of depth-dependent collimator blurring in SPECT. In future work we plan to investigate the validity of the approximations introduced in the analysis, and to formulate the method for more general models of the data collection process. The design of a robust inverse filter for noisy data will be crucial for the successful application of the method to measured data in SPECT.

7. ACKNOWLEDGMENTS

We are grateful to Drs. R.H.T. Bates, W.G. Hawkins, G.T. Herman, A. Markoe and G. Muehllehner for useful discussions. This work is supported by grant HL-28438 from the National Institutes of Health.

8. REFERENCES

1. B. Y. Croft, *Single-Photon Emission Computed Tomography*, Year Book Medical Publishers, Inc., Chicago (1986).
2. S. L. Heller and P. N. Goodwin, "SPECT instrumentation: performance, lesion detection, and recent innovations," *Seminars in Nuclear Medicine* 17(3), 184-199 (1987).
3. G. Muehlelehner, "Scintillation camera collimators," in *Imaging for Medicine*, vol. 1, S. Nudelman and D. D. Patton, eds., chap. 2, pp. 77-87, Plenum Publishing Corp., New York (1980).
4. J. A. Sorenson and M. E. Phelps, *Physics in Nuclear Medicine*, Second Edition, Grune & Stratton, Inc., Orlando, Florida (1987).
5. M. T. Madsen and C. H. Park, "Enhancement of SPECT images by Fourier filtering the projection image set," *The Journal of Nuclear Medicine*, 26(4), 395-402, (1985).
6. P. R. Edholm, R. M. Lewitt and B. Lindholm, "Novel properties of the Fourier decomposition of the sinogram," in *International Workshop on Physics and Engineering of Computerized Multidimensional Imaging and Processing*, Proceedings of SPIE, 671, 8-18 (1986).
7. K. Ogawa, S. Paek, M. Nakajima, S. Yuta, A. Kubo and S. Hashimoto, "Correction of collimator aperture using a shift-variant deconvolution filter in gamma camera emission CT," in *Medical Imaging II: Image Formation, Detection, Processing, and Interpretation*, Proceedings of SPIE, 914, 699-706 (1988).
8. W. G. Hawkins, P. K. Leichner and N.-C. Yang, "The circular harmonic transform for SPECT reconstruction and boundary conditions on the Fourier transform of the sinogram," *IEEE Transactions on Medical Imaging*, 7(2), 135-148 (1988).
9. R. M. Lewitt, "Reconstruction algorithms: transform methods," *Proceedings of the IEEE*, 71(3), 390-408 (1983).
10. Y. Censor, "Finite series-expansion reconstruction methods," *Proceedings of the IEEE*, 71(3), 409-419 (1983).
11. G. N. Watson, *A Treatise on the Theory of Bessel Functions*, Second Edition, Cambridge University Press, New York (1944).
12. F. Natterer, *The Mathematics of Computerized Tomography*, John Wiley & Sons, Chichester, England (1986).
13. M. Abramowitz and I. A. Stegun, Eds., *Handbook of Mathematical Functions with Formulas, Graphs and Mathematical Tables*, Dover Publications, Inc., New York, 9th printing (1965).
14. J. T. Payne, L. E. Williams, R. A. Ponto, M. E. Goldberg and M. K. Loken, "Comparison and performance of Anger cameras," *Radiology*, 109(2), 381-386, (1973).
15. D. R. Gilland, B. M. W. Tsui, W. H. McCartney, J. R. Perry and J. Berg, "Determination of the optimum filter function for SPECT imaging," *The Journal of Nuclear Medicine*, 29(5), 643-650 (1988).
16. M. A. King, P. W. Doherty, R. B. Schwinger, D. A. Jacobs, R. E. Kidder and T. R. Miller, "Fast count-dependent digital filtering of nuclear medicine images: concise communication," *The Journal of Nuclear Medicine*, 24(11), 1039-1045 (1983).
17. M. A. King, P. W. Doherty and R. B. Schwinger, "A Wiener filter for nuclear medicine images," *Medical Physics*, 10(6), 876-880 (1983).
18. M. A. King, R. B. Schwinger, P. W. Doherty and B. C. Penney, "Two-dimensional filtering of SPECT images using the Metz and Wiener filters," *The Journal of Nuclear Medicine*, 25(11), 1234-1240, (1984).
19. M. A. King, S. J. Glick, B. C. Penney and R. B. Schwinger, "Interactive visual optimization of SPECT prereconstruction filtering," *The Journal of Nuclear Medicine*, 28(7), 1192-1198 (1987).




Fig.1 Projection Geometry




Fig.2 Geometry of collimator response function for hole centered on line (s, θ) and point source at (s', t) .




Fig.3 A point object in both image (x, y) and detector (s, t) coordinates




Fig.4 Typical plots of $\Phi(\theta)$ and $\cos(\Phi(\theta))$ for a fixed (ω, n) .




Fig.5 Comparison of the sinograms from an ideal detector (a) and a collimated detector (b), and corresponding reconstructed images from those sinograms, also the maps of Fourier coefficients of the sinograms.




Fig.6 Contribution of source points to the map of Fourier coefficients of the sinogram.




Fig.7 Demonstration of the simple simulation to modify SPECT sinogram and reconstruction. (a) is the collimated data. (b) shows the data modified by discarding Fourier coefficients which are corresponding to the part of projection data contributed by sources farther away from the detector than the center of rotation. (c) is the complementary part of the data of (b). It shows the blurring of collimator at large distances.




Fig.8 Simulation of the data acquired from a collimated detector system. High sensitivity parallel hole collimator with a bore depth to diameter ratio of 5:1 is simulated. Attenuation and septal penetration are ignored in the simulation. Deblurring filter is based on Gaussian shaped PSF assumption. Phantom is composed of circles with different sizes within an unit image area. Image size is 135 by 135. Blurred sinogram is double Fourier decomposed and processed in Fourier domain, followed by an inverse Fourier transform to get a modified sinogram. The modified sinogram leads to a deblurred image reconstruction.

- 1-A** Fourier decomposition of the sinogram from an ideal detector system with distance invariant sensitivity.
- 1-B** Reconstruction from the ideal sinogram by standard filtered back-projection algorithm.
- 2-A** Fourier decomposition of the blurred sinogram from the collimated detector system.
- 2-B** Reconstruction resulting from the blurred sinogram.
- 3-A** Fourier decomposition in 2-A processed by the deblurring filter.
- 3-B** Reconstruction resulting from the modified sinogram obtained by inverse transform of 3-A.